

## DAFTAR PUSTAKA

- [1] E. Vañó *et al.*, *ICRP135 Annals of the ICRP*, vol. 44, no. 1. 2017. [Online]. Available: [www.icrp.org](http://www.icrp.org)
- [2] E. Hiswara and D. Kartikasari, "Dosis Pasien Pada Pemeriksaan Rutin Sinar-X Radiologi Diagnostik," *J. Sains dan Teknol. Nukl. Indones.*, vol. 16, no. 2, p. 71, 2015, doi: 10.17146/jstni.2015.16.2.2359.
- [3] S. Rekaman Dokumen, "Pedoman Teknis Penyusunan Tingkat Panduan Diagnostik Atau Diagnostic Reference Level (Drl) Nasional," *Jakarta*, no. 8, p. 63858275, 1012.
- [4] Vienna, "Technical Reports Series No. 457 IAEA," *Dosim. DIAGNOSTIC Radiol. AN Int. CODE Pract.*, vol. 457, 2007.
- [5] D. R. Dance, C. L. Skinner, K. C. Young, J. R. Beckett, and C. J. Kotre, "Additional factors for the estimation of mean glandular breast dose using the UK mammography dosimetry protocol," *Phys. Med. Biol.*, vol. 45, no. 11, pp. 3225–3240, 2000, doi: 10.1088/0031-9155/45/11/308.
- [6] C. Songsaeng, K. Mommontani, and K. Panatbul, "Original Article An education for breast cancer prevention with a digital breast X-ray machine at Sanuit Naivech , Songkhla Nakhon Nithon Hospital Local diagnostic reference levels for breast screening using digital mammography at Tanyawej Breast Center," *Local diagnostic Ref. Lev. breast Screen. using Digit. Mammogr. Tanyawej Breast Cancer*, vol. 44, no. 1, pp. 1–8, 2019.
- [7] M. Talbi, M. El Mansouri, O. Nhila, Z. Tahiri, K. Eddaoui, and M. Khalis, "Local diagnostic reference levels (LDRLs) for full-field digital mammography (FFDM) and digital breast tomosynthesis (DBT) procedures in Morocco," *J. Med. Imaging Radiat. Sci.*, Mar. 2022, doi: 10.1016/J.JMIR.2022.03.008.
- [8] I. Garba *et al.*, "Local diagnostic reference levels for digital mammography : Two hospitals study in northwest , Nigeria Local diagnostic reference levels for digital mammography : Two hospitals study in northwest , Nigeria," *J.*

- Med. Imaging Radiat. Sci.*, no. June, 2021, doi: 10.1016/j.jmir.2021.03.035.
- [9] BAPETEN, “Penetapan Nilai Tingkat Panduan Diagnostik Indonesia (Indonesian Diagnostic Reference Level) Untuk Modalitas Sinar-X Ct Scan Dan Radiografi Umum,” *Angew. Chemie Int. Ed.* 6(11), 951–952., pp. 2013–2015, 2021.
- [10] E. Vañó *et al.*, “ICRP 135 - Diagnostic Reference Levels in Medical Imaging,” *Protection, International Commission on Radiological*, vol. 44, no. 1. pp. 1–143, 2017. [Online]. Available: [www.icrp.org](http://www.icrp.org)
- [11] BAPETEN, “Pedoman Teknis Penerapan Tingkat Panduan Diagnostik Indonesia (Indonesian Diagnostic Reference Level),” no. 8, pp. 1–54, 2021.
- [12] M. Drl, “Pengukuran Dosis Kelenjar Rata-Rata di Tiga Pusat Mammografi di Kashaan : Suatu Abstrak,” vol. 4, no. 4, pp. 285–291, 2021.
- [13] D. R. Dance, C. L. Skinner, K. C. Young, J. R. Beckett, and C. J. Kotre, “Dance, Additional factors for the estimation of mean glandular breast dose using the UK mammography dosimetry protocol,” *Phys. Med. Biol.*, vol. 45, no. 11, pp. 3225–3240, 2000.
- [14] E. moayaad; C. P. P. J. L. F. Mark, “DRL for Digital Mamography, time for a new paradigm.” The University of sydney, sydney, p. 10.
- [15] E. M. L. Jerrold T. Bushberg, J. Anthony Seibert, “the-essential-physics-for-medical-imaging-2nd Edition.” LIPPINCOTT WILLIAM & WILKINS, California.
- [16] J. M. Jerrold T. Bushberg, J. Anthony Seibert, Edwin M. Leidholdt Jr., “The Essential Physics Of Medical Imaging.” <https://archive.org/details/JerroldT.BushbergJ.AnthonySeibertEdwinM.LeidholdtJr.JohnM.BooneTheEssentialPhysi/page/n251/mode/2up?view=theater> (accessed May 13, 2022).
- [17] G. J. A. J, “Grids Improve Mammography Contrast | Semantic Scholar,” 2004. <https://www.semanticscholar.org/paper/Grids-Improve-Mammography-Contrast-Gray-Princehorn/cc1bd5729a2d121db3e3cab1b2fc3eed0aa1a3a> (accessed May 31, 2022).
- [18] L. C. Miller, “Rethinking Mammography Principles : What works and why,” *Breast Imaging Symp.*, 2016.

- [19] “What Does It Mean to Have Dense Breasts? | CDC.”  
[https://www.cdc.gov/cancer/breast/basic\\_info/dense-breasts.htm](https://www.cdc.gov/cancer/breast/basic_info/dense-breasts.htm) (accessed May 13, 2022).
- [20] “What’s the best way to screen dense breasts?”  
<https://www.fredhutch.org/en/news/center-news/2020/02/what-s-the-best-way-to-screen-dense-breasts-.html> (accessed May 13, 2022).
- [21] C. Songsaeng, “Journal of Associated Medical Sciences,” no. November 2021, 2022, doi: 10.12982/JAMS.2022.002.
- [22] E. Dzidzornu, S. K. Angmorte, B. B. Ofori-manteaw, S. Aboagye, K. Dze, and E. K. Ofori, “Radiography Mammography Diagnostic Reference Levels ( DRLs ) in Ghana,” no. xxxx, 2020, doi: 10.1016/j.radi.2020.11.022.
- [23] M. T. Bahreyni Toossi, H. Zare, S. Bayani Roodi, M. Hashemi, F. Akbari, and M. Malekzadeh, “Towards proposition of a diagnostic reference level for mammographic examination in the greater khorasan province, Iran,” *Radiat. Prot. Dosimetry*, vol. 155, no. 1, pp. 96–99, 2013, doi: 10.1093/rpd/ncs317.
- [24] M. E. Suleiman, M. F. McEntee, L. Cartwright, J. Diffey, and P. C. Brennan, “Diagnostic reference levels for digital mammography in New South Wales,” *J. Med. Imaging Radiat. Oncol.*, vol. 61, no. 1, pp. 48–57, 2017, doi: 10.1111/1754-9485.12540.
- [25] A. Aliasgharzadeh, H. Moradi, T. Talakesh, E. Motallebzadeh, G. Ataei, and M. Haghighi, “Mean Glandular Dose Measurement in Three Mammography Centers in Kashan : An approach to Provide a Local DRL,” vol. 8, no. 4, pp. 285–291, 2021.
- [26] C. Songsaeng and U. P. Songkla, “Jurnal Ilmu Kedokteran Terkait,” no. November 2021, 2022, doi: 10.12982/JAMS.2022.002.
- [27] N. Mohd Norsuddin, S. Segar, R. Ravintaran, N. Mohd Zain, and M. K. Abdul Karim, “Local Diagnostic Reference Levels for Full-Field Digital Mammography and Digital Breast Tomosynthesis in a Tertiary Hospital in Malaysia,” *Healthc.*, vol. 10, no. 10, 2022, doi: 10.3390/healthcare10101917.
- [28] O. Ciraj-Bjelac, S. Beciric, D. Arandjic, D. Kosutic, and M. Kovacevic, “Mammography radiation dose: Initial results from Serbia based on mean glandular dose assessment for phantoms and patients,” *Radiat. Prot.*

*Dosimetry*, vol. 140, no. 1, pp. 75–80, 2010, doi: 10.1093/rpd/ncq040.



## LAMPIRAN

LOGBOOK PENYINARAN UNTUK TIAP JENIS PEMERIKSAAN DENGAN MAMOGRAFI

JAKARTA SELATAN								INDIKATOR DOSIS	
NO	INSIAL PASIEN	UMUR (Tahun)	GENDER (M/F)	PROYEKSI	kVp	mAs	CBT (mm)	ESAK (mGy)	AGD (mGy)
1	J.E	50	F	RCC	28	96	27	3,78	1,93
				LCC	29	98	30	3,74	1,76
				RMLO	28	97	28	3,8	1,83
				LMLO	29	98	30	3,72	1,72
2.	R.Y	50	F	RCC	29	139	42	6,27	2,49
				LCC	29	145	42	6,54	2,4
				RMLO	29	138	43	6,26	2,48
				LMLO	29	142	42	6,51	2,37
3	M.G	43	F	RCC	29	139	38	6,18	2,38
				LCC	29	131	35	5,77	2,36
				RMLO	29	138	38	6,18	2,38
				LMLO	29	131	36	5,82	2,36
4	A.T	45	F	RCC	29	156	49	7,2	2,36
				LCC	29	155	48	7,13	2,37
				RMLO	29	156	49	7,22	2,37
				LMLO	29	154	48	7,11	2,36
5	A.K	63	F	RCC	29	118	47	5,41	1,96
				LCC	29	131	49	6,05	2,14
				RMLO	29	119	48	5,45	1,99
				LMLO	29	131	49	6,02	2,16
6	W.D	53	F	RCC	30	227	59	11,73	3,62
				LCC	30	225	61	11,71	3,53
				RMLO	30	229	60	11,73	3,6
				LMLO	30	223	61	11,71	3,53
7	Y.V	45	F	RCC	29	144	51	6,69	2,16
				LCC	29	164	58	7,81	2,3
				RMLO	29	148	54	6,59	2,19
				LMLO	29	156	58	7,67	2,32
8	D.H	44	F	RCC	30	230	64	12,9	3,33
				LCC	30	235	67	12,48	3,33
				RMLO	30	229	62	12,87	3,33
				LMLO	30	232	69	12,38	3,32
9	N.T	49	F	RCC	29	139	54	6,53	2,03
				LCC	29	188	62	9,08	2,53
				RMLO	29	142	58	6,3	2,22
				LMLO	29	187	61	9,2	2,64
10	D.T	45	F	RCC	29	110	41	4,94	1,85
				LCC	29	112	41	5,03	1,88
				RMLO	29	106	42	4,86	1,92
				LMLO	29	119	41	5,12	1,83
11	L.U	45	F	RCC	29	162	46	7,4	2,51
				LCC	29	156	50	7,23	2,34
				RMLO	29	163	48	7,39	2,49
				LMLO	29	145	52	7,29	2,76
12	S.S	50	F	RCC	29	108	32	4,71	2,14
				LCC	29	108	32	4,71	2,14
				RMLO	29	108	31	4,71	2,13
				LMLO	29	109	32	4,71	2,14
13	L.H	53	F	RCC	29	202	58	9,62	1,74
				LCC	30	200	59	10,33	1,87
				RMLO	29	210	58	9,59	1,79
				LMLO	30	200	61	10,43	1,78
14	A.U	44	F	RCC	30	212	63	11,11	3,1
				LCC	30	205	66	10,35	2,94
				RMLO	30	210	60	11,21	3,33
				LMLO	30	208	69	10,61	2,89
15	S.H	60	F	RCC	29	136	54	6,39	2,12
				LCC	29	137	53	6,41	2,16
				RMLO	29	135	54	6,39	2,12
				LMLO	29	137	53	6,41	2,14

16	H.S	70	F	RCC	29	170	46	7,77	2,78
				LCC	29	140	42	6,31	2,41
				RMLO	29	172	45	7,67	2,87
				LMLO	29	140	42	6,3	2,41
17	K.D	68	F	RCC	29	150	51	6,97	2,39
				LCC	39	155	50	7,18	2,48
				RMLO	29	149	51	7,01	2,36
				LMLO	39	155	50	7,22	2,48
18	P.M	53	F	RCC	29	160	56	7,85	2,52
				LCC	29	166	59	7,64	2,37
				RMLO	29	164	52	7,81	2,78
				LMLO	29	171	59	7,61	2,39
19	A.N	47	F	RCC	29	192	52	8,96	2,81
				LCC	29	190	51	8,83	2,8
				RMLO	29	196	54	8,89	2,83
				LMLO	29	192	51	8,85	2,91
20	L.N	51	F	RCC	29	121	47	5,55	2,01
				LCC	29	129	49	5,96	2,1
				RMLO	29	124	48	5,45	2,01
				LMLO	29	129	49	5,89	2,18
21	T.K	48	F	RCC	30	205	54	10,41	3,22
				LCC	30	220	59	11,37	3,3
				RMLO	30	208	55	10,38	3,42
				LMLO	30	219	60	11,63	3,31
22	S.F	45	F	RCC	29	134	38	5,96	2,31
				LCC	29	132	38	5,87	2,28
				RMLO	29	134	36	6,02	2,42
				LMLO	29	134	38	5,43	2,32
23	J.N	64	F	RCC	28	103	39	4,22	1,7
				LCC	28	97	36	3,94	1,68
				RMLO	28	104	38	4,12	1,53
				LMLO	28	97	36	3,89	1,73
24	T.J	64	F	RCC	29	136	40	6,09	2,38
				LCC	29	99	41	4,45	1,76
				RMLO	29	123	46	6,21	2,11
				LMLO	29	111	42	4,84	1,56
25	F.E	60	F	RCC	28	95	32	3,8	1,74
				LCC	29	100	37	4,43	1,87
				RMLO	28	98	34	3,63	1,56
				LMLO	29	102	36	4,73	1,79



## LOGBOOK PENYINARAN UNTUK TIAP JENIS PEMERIKSAAN DENGAN MAMOGRAFI

JAKARTA PUSAT								INDIKATOR DOSIS	
NO	INSIAL PASIEN	UMUR (Tahun)	GENDER (M/F)	PROYEKSI	kVp	mAs	CBT (mm)	ESAK (mGy)	AGD (mGy)
1	J.A	52	F	RCC	29	122	53	5,71	1,94
				LCC	29	137	54	6,43	2,14
				RMLO	29	129	53	5,73	1,95
2	H.N	49	F	LMLO	29	132	52	6,34	2,1
				RCC	29	139	53	6,51	2,05
				LCC	29	125	52	5,83	1,87
3	K.S	56	F	RMLO	29	142	54	6,34	2,22
				LCC	29	123	65	6,01	1,64
				LMLO	29	129	55	6,43	1,98
4	A.W	60	F	RCC	29	123	62	5,29	1,69
				LCC	29	123	58	6,32	1,69
				RMLO	29	129	62	5,63	1,59
5	P.B	56	F	RCC	29	129	62	6,23	1,89
				LCC	29	118	58	5,41	1,8
				RMLO	29	142	80	6,56	2,04
6	S.C	46	F	LMLO	29	103	80	12,35	1,97
				RCC	30	237	80	13,18	3,23
				LCC	30	222	80	12,35	3,03
7	I.F	49	F	RMLO	30	265	78	13,67	2,89
				LCC	30	201	80	12,22	2,98
				LMLO	30	201	80	12,22	2,98
8	N.R	72	F	RCC	28	89	44	3,71	1,34
				LCC	29	139	52	6,48	2,08
				RMLO	29	92	64	2,76	1,76
9	A.F	47	F	LMLO	28	142	67	6,43	1,05
				RCC	29	101	44	4,58	1,65
				LCC	28	110	40	4,52	1,69
10	N.H	51	F	RMLO	29	119	29	4,25	1,71
				LCC	27	121	27	4,43	1,66
				RCC	28	88	46	3,69	1,38
11	Y.S	37	F	LCC	28	107	53	4,6	1,65
				RMLO	28	92	49	3,89	1,23
				LMLO	28	121	51	4,21	1,76
12	M.C	49	F	RCC	27	80	45	3,05	0,95
				LCC	27	80	46	3,06	0,95
				RMLO	27	81	44	3,1	1,01
13	S.D	41	F	LMLO	27	79	46	3,03	0,98
				RCC	29	108	51	5,09	1,7
				LCC	29	188	55	9,11	2,68
14	S.M	58	F	RMLO	29	112	63	6,12	1,92
				LCC	29	100	50	4,27	1,27
				LMLO	29	202	59	11,12	2,48
15	M.L	54	F	RCC	29	101	42	4,55	1,69
				LCC	29	99	39	4,42	1,71
				RMLO	29	111	49	4,89	1,71
16	M.L	54	F	LMLO	29	119	43	4,36	1,89
				RCC	29	190	70	9,43	2,43
				LCC	29	156	63	7,56	2,11
17	M.L	54	F	RMLO	29	211	78	9,08	2,56
				LCC	29	167	71	6,78	1,98
				LMLO	29	167	71	6,78	1,98
18	M.L	54	F	RCC	28	100	51	4,27	1,27
				LCC	28	100	50	4,25	1,28
				RMLO	28	100	56	4,56	1,37
19	M.L	54	F	LMLO	28	103	52	4,29	1,34
				RCC	28	72	29	2,66	1,33
				LCC	27	65	27	2,07	1,11
20	M.L	54	F	RMLO	28	72	29	2,87	1,45
				LCC	27	65	28	2,3	1,28
				LMLO	27	65	28	2,3	1,28
21	M.L	54	F	RCC	28	108	48	4,56	1,64
				LCC	28	112	46	4,7	1,71
				RMLO	28	108	48	4,67	1,82
22	M.L	54	F	LMLO	28	112	46	4,32	1,67

16	L.J	67	F	RCC	29	112	46	4,7	1,71
				LCC	28	111	56	5,25	1,73
				RMLO	28	109	53	4,97	1,87
				LMLO	29	116	61	5,82	2,02
17	P.D	51	F	RCC	29	205	79	10,51	2,58
				LCC	30	222	87	12,66	2,9
				RMLO	29	212	92	11,32	3,1
				LMLO	30	262	89	12,89	2,69
18	J.C	55	F	RCC	29	155	69	7,67	2,13
				LCC	29	129	64	6,27	1,87
				RMLO	29	162	71	7,45	2,56
				LMLO	29	113	68	6,88	1,98
19	M.A.G	69	F	RCC	28	112	51	4,78	1,66
				LCC	28	102	50	4,34	1,53
				RMLO	28	132	52	4,89	1,49
				LMLO	28	110	50	4,65	1,72
20	R.C	50	F	RCC	28	70	27	2,77	1,46
				LCC	28	76	28	3,04	1,54
				RMLO	28	73	28	3,12	1,56
				LMLO	28	87	29	2,56	1,62
21	Z.F	56	F	RCC	30	207	71	11,15	3,03
				LCC	30	184	69	9,84	2,75
				RMLO	30	215	76	12,14	3,33
				LMLO	30	196	62	10,22	3,01
22	R.T	45	F	RCC	30	247	63	12,94	3,59
				LCC	30	227	62	11,85	3,33
				RMLO	30	251	62	12,3	3,4
				LMLO	30	210	62	11,67	4,06
23	R.I	38	F	RCC	30	230	69	12,3	3,22
				LCC	30	198	67	10,52	2,83
				RMLO	30	243	69	11,57	2,93
				LMLO	30	208	71	10,54	3,11
24	K.W	65	F	RCC	29	137	52	6,39	2,18
				LCC	29	117	48	5,38	1,94
				RMLO	29	142	51	6,23	1,83
				LMLO	29	111	53	5,45	2,01
25	N.J	56	F	RCC	30	196	62	10,23	2,94
				LCC	30	196	63	10,27	2,92
				RMLO	30	196	61	11,32	2,98
				LMLO	30	195	66	11,45	3,26